

**CRB Analyse Service GmbH**  
Dr S Pierdzig  
Bahnhofstrasse 14  
D-37181 Hardegsen

**37181**

**GERMANY**

## **AIMS Round 46 Report**

Please find enclosed your results for AIMS Round 46.

Regards


Delia Lomas-Fletcher  
Senior Scientist  
AIMS PT Manager

Tel: 01298 218553


Fax: 01298 218172

Email: [proficiency.testing@hsl.gov.uk](mailto:proficiency.testing@hsl.gov.uk)

Distribution of Individual Scores for this Round




Distribution of 3-Round Cumulative Scores



Quantitative Results for Sample 1

Histogram of Quantitative Results in Round 46



# Asbestos in Materials Individual Results Round 46

for Laboratory Number 1640

Printed on 15 Mar 2012

Your Performance Score:

**32**

Your Performance Classification:

**Satisfactory**

Your Score for this round:

**20**

Sample 1	Your Results	Assigned Results	Median	Your Score
	Other	No Asbestos		0
	No Asbestos			

Sample 2	Your Results	Assigned Results	Median	Your Score
	Amosite	Amosite		0
	Chrysotile	Chrysotile		

Sample 3	Your Results	Assigned Results	Median	Your Score
	No Asbestos	Tremolite		20

Sample 4	Your Results	Assigned Results	Median	Your Score
	Amosite	Amosite		0
	MMMF			

